

## maXs: Microcalorimeter Arrays for High-Resolution X-ray Spectroscopy atGSI/FAIR

C. Pies<sup>a</sup>; S. Schäfer<sup>a</sup> · S. Heuser<sup>a</sup> ·  
S. Kempf<sup>a</sup>, A. Pabinger<sup>a</sup>, J.-P. Porsta<sup>a, b</sup>,  
P. Ranitsch<sup>a</sup>, N. Foerster<sup>a</sup>, D. Hengstler<sup>a</sup>,  
A. Kampkötter<sup>a</sup>, L. Gastaldo<sup>a</sup>, T. Wolf<sup>a</sup>,  
A. Fleischmann<sup>a</sup>, C. Enss<sup>a</sup>

**Abstract** - Highly-charged heavy ions like U91+ provide unique conditions for the investigation of relativistic and quantum electrodynamical effects in strong electromagnetic fields. We present two x-ray detectors developed for high-resolution spectroscopy on highly-charged heavy ions. Both detectors consist of metallic magnetic calorimeters (MMCs) forming linear eight-pixel arrays. The first detector, maXs-20, is developed for the detection of x-rays up to 20 keV with an energy resolution below 3 eV. The second device, maXs-200, is designed for x-ray energies up to 200 keV with an energy resolution of 40 eV. The results of characterization measurements of both detectors will be shown and discussed. Furthermore, we present a prototype MMC for soft x-rays with improved magnetic flux coupling. In first characterization measurements the energy resolution of this device was 2.0 eV (FWHM) for x-rays up to 6 keV.

**Keywords** - MMC, metallic magnetic calorimeters, low-temperature detectors, x-ray detectors, QED effects in ions

Submitted to ESNF November 30, 2011; accepted December 7, 2011.

Reference ST288, Category 4

Published in Journal of Low Temperature Physics 167, Numbers 3-4, 269-279 , (2012).