

Voltage Metrology with Superconductive Electronics

Sam P. Benz

National Institute of Standards and Technology, Boulder, Colorado, 80305 USA
E-mail: benz@nist.gov

Abstract - Josephson's discovery in 1962 of the quantum behavior of superconducting junctions enabled a revolution in precision voltage measurement that replaced electrochemical cells, which are artifact standards whose behavior depends upon environmental conditions, with quantum-based standards, whose values are intrinsically accurate and can be reproduced anywhere. Many technological advances in junction fabrication, superconducting integrated circuit technology, bias techniques, and instrumentation were required to achieve the present generation of practical ac and dc voltage standard systems. Quantum-based 10 V programmable Josephson voltage standards and 2 V rms Josephson arbitrary waveform synthesizers are now used in a wide range of metrology applications, calibration laboratories and precision measurement experiments. For metrology, these systems are used for measuring dc and ac voltage, ac power, and impedance. They are also key instruments in precision measurement experiments of mass and temperature to determine more accurate values of the Planck and Boltzmann constants. I will review major technological advances with a focus on the superconducting devices and circuits and describe the current state-of-the-art research and development in superconductive analog and digital circuits that may lead to improved precision measurement of voltage and low-distortion signals for rf communications.

References

1. B. D. Josephson, "Possible new effects in superconductive tunneling," *Phys. Lett.*, vol. 1, pp. 251-253, July 1962.
2. J. Niemeyer, "Josephson voltage standards," in Handbook of Applied Superconductivity, vol. 2, B. Seeber, ed., Institute of Physics, Philadelphia, PA, 1998, pp. 1813-34.
3. S. P. Benz, "Synthesizing accurate voltages with superconducting quantum-based standards," *IEEE Instr. Meas. Magazine*, June 2010, pp. 8-13.
4. H. Yamamori, *et al.*, "A 10V programmable Josephson voltage standard circuit with a maximum voltage of 20 V", *Supercond. Sci. Technol.*, vol. 21, 105007-6, July 2008, DOI:10.1088/0953-2048/21/10/105007.
5. F. Müller, *et al.*, "NbSi Barrier Junctions Tuned for Metrological Applications up to 70 GHz: 20 V Arrays for Programmable Josephson Voltage Standards," *IEEE Trans. Appl. Supercond.*, vol. 24, pp. 1101005-5, Feb. 2014, DOI: 10.1109/TASC.2013.2284834.
6. A. Rüfenacht, *et al.*, "Cryocooled 10 V programmable Josephson voltage standard," *IEEE Trans. Inst. Meas.*, vol. 64, pp. 1477-1482, June 2015, DOI: 10.1109/TIM.2014.2374697.
7. N. E. Flowers-Jacobs, *et al.*, "Two-volt Josephson arbitrary waveform synthesizer using Wilkinson dividers," *IEEE Trans. Appl. Supercond.*, in press, DOI: 10.1109/TASC.2016.2532798.
8. Jifeng Qu, *et al.*, "Improved electronic measurement of the Boltzmann constant by Johnson noise Thermometry," *Metrologia*, vol. 52, pp. 441-451, June 2015, DOI: 10.1088/0026-1394/50/5/441
9. Urano, *et al.*, "Johnson noise thermometry based on Integrated quantum voltage noise source," *IEEE Trans. Appl. Supercond.*, vol. 26, pp. 1800305-5, April 2016, DOI: 10.1109/TASC.2016.2524472.

Acknowledgements

I thank and recognize all current and former members and guest researchers of the NIST Quantum Voltage Project and collaborators and colleagues from national and international measurement institutes and universities, all of whom have contributed to this field of research.

Keywords (Index Terms) – Metrology, standard, voltage standard, quantum, superconductive, electronics, integrated circuit.

IEEE CSC & ESAS SUPERCONDUCTIVITY NEWS FORUM (Global Edition), No. 38, October 2016.
This ASC 2016 plenary (annotated) presentation was not submitted to *IEEE Trans, Appl. Supercond.*

IEEE/CSC & ESAS SUPERCONDUCTIVITY NEWS FORUM (global edition), No. 38, October 2016.
Annotated version received February 23, 2017; selected March 08, 2017. Reference RP76; Category 4.
This plenary oral presentation 5PL 01 was given at ASC 2016 and initially published by SNF without notes.
Auxiliary video is accessible via a link in notes. No manuscript was submitted to *IEEE Trans. Appl. Supercond.*